

ICS 17.240
CCS A 58

JJG

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

JJG 912-2010

Replacing JJG 912-1996

**Verification Regulation of Ionization Chamber Dosimeters Used
In Radiotherapy
(English Translation)**

Issued on: June 10, 2010

Implemented on: December 10, 2010

**Issued by: General Administration of Quality Supervision, Inspection and
Quarantine**

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to therapeutic levels of ionization chamber dosimeters initial verification, testing and use of follow-up inspection. HVL test range is 0.027mmAl to 3.00mmCu of low-energy X-rays and ^{60}Co or ^{137}Cs gamma ray radiation field proceed.

Document data

International classification (ICS): 17.240.

Chinese classification (CCS): A58.

Estimated length of the Chinese original: 15,157 words.

Supersedes the earlier edition JJG 912-1996.

Standards quoted by this document

IEC 60731; JJF 1001-1998; JJF 1059-1999; JJF 1035-2006; GB/T 17857-1999