

INDUSTRY STANDARD OF THE PEOPLE'S REPUBLIC OF CHINA

JJG 969-2002

Verification regulation of gamma radioimmunoassay counters (English Translation)

Issued on: April 15, 2002

Implemented on: July 1, 2002

Issued by: The General Administration of Quality Supervision, Inspection
and Quarantine issued, This regulation was approved by the General
Administration of Quality Supervision, Inspection and Quarantine on
April 15, 2002, and

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard applies to single-and multi-sensor probe ?? RIA counter initial verification, testing and use of the follow-up inspection. Count rate in the range of 3000/min ~ 200000/min, used radionuclide as ^{125}I .

Document data

Chinese classification (CCS): A58.

Estimated length of the Chinese original: 10,161 words.

Standards quoted by this document

GB 10255-1996; GB 4960-1996; GB/T 4080-1983; JJG 44-1999