

ICS 11.020
CCS C 08

SN

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

SN/T 2745-2010

**Test method for non-destructive detection of leaks in medical
packaging trays. CO2 tracer gas method**

医用包装盘泄漏性无损检测试验方法 二氧化碳法

(English Translation)

Issued on: November 1, 2010

Implemented on: May 1, 2011

Issued by: General Administration of Customs

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the method of carbon dioxide tracer gas leak of medical packaging tray for non-destructive testing of test methods. This method is only suitable for detecting non-sealed and empty plate in the production of medical packaging design and manufacturing process produces a minimum diameter of 50µm pinhole leaks and cracks in the same size.

Document data

International classification (ICS): 11.020.

Chinese classification (CCS): C08.

Estimated length of the Chinese original: 9,943 words.

Adopted international standard: ASTM F2227-2007, IDT.

Standards quoted by this document

ASTM D996; ASTM F1327