

ICS 11.040
CCS C 30

SN

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

SN/T 3062.3-2011

**Packaging materials for terminally sterilized medical devices for
import. Part 3: Sealable pouches and reels of porous materials
and plastic film construction. Requirements and test methods**

进口医疗器械灭菌包装 第3部分：透气材料与塑料膜组成的可密封组合袋和卷材
要求和试验方法

(English Translation)

Issued on: September 9, 2011

Implemented on: April 1, 2012

Issued by: General Administration of Customs

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard provides for sterile medical device packaging sealed bags and roll test methods and requirements. The main purpose of this section is to establish a sterile barrier system or packaging to ensure that appropriate standards sterilized medical devices sterile. This standard specifies that all devices are disposable.

Document data

International classification (ICS): 11.040.

Chinese classification (CCS): C30.

Estimated length of the Chinese original: 10,158 words.

Adopted international standard: BS EN 868-5-2009, IDT.

Standards quoted by this document

ASTM D882-1995; EN 868-2; EN 868-3; EN 868-6; EN 868-7; EN 868-9; EN 868-10; EN ISO 11140-1; EN ISO 11607-1-2006; ISO 8601