

ICS 13.280
CCS C 57

WS

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

WS 818-2023

**Standard for testing of quality control in X-ray cone beam
computed tomography equipment
(English Translation)**

Issued on: March 7, 2023

Implemented on: March 1, 2024

Issued by: National Health Commission

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the general requirements, testing items, testing methods and technical requirements for the quality control testing of cone beam X-ray computed tomography (CCT) equipment. This standard applies to the quality control testing of CCT equipment, including oral CBCT equipment, C-arm angiography machine with CBCT function, mammary vein CCT equipment and CCT equipment in electron linear accelerators: CBCT equipment in proton heavy ion radiation therapy shall be implemented with reference .

Document data

International classification (ICS): 13.280.

Chinese classification (CCS): C57.

Estimated length of the Chinese original: 28,299 words.