

ICS 11.020
CCS C 50

WS

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

WS/T 408-2024

Replacing WS/T 408-2012; WS/T 416-2013; WS/T 420-2013; WS/T 492-2016

**Guideline for the verification of analytical performance of
quantitative examination procedures
(English Translation)**

Issued on: May 9, 2024

Implemented on: November 1, 2024

Issued by: National Health Commission

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the principles and methods for verification of the analytical performance (precision, accuracy, linearity, and specificity) of quantitative test procedures. This standard applies to the verification of the analytical performance of quantitative test procedures by clinical laboratories of medical institutions. This standard does not apply to the establishment or confirmation of the analytical performance of test procedures.

Document data

International classification (ICS): 11.020.

Chinese classification (CCS): C50.

Estimated length of the Chinese original: 15,157 words.

Supersedes the earlier edition WS/T 408-2012;WS/T 416-2013;WS/T 420-2013;WS/T 492-2016.