

ICS 73.080
CCS D 52

YB

**INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

YB/T 4583.1-2017

**Mulite. Determination of silicon, ferrous, calcium, magnesium,
manganese and phosphorous content. Inductively coupled
plasma atomic emission spectrometry**

莫来石二氧化硅、三氧化二铁、氧化钙、氧化镁、二氧化钛和五氧化二磷含量的
测定电感耦合等离子体原子发射光谱法

(English Translation)

Issued on: April 12, 2017

Implemented on: October 1, 2017

Issued by: Ministry of Industry and Information Technology

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the use of inductively coupled plasma atomic emission spectrometry (ICP-AES) determination of SiO₂, Fe₂O₃, CaO, MgO, TiO₂ and P₂O₅ content. This section applies to mullite SiO₂, Fe₂O₃, CaO, MgO, TiO₂ and P₂O₅ content determination.

Document data

International classification (ICS): 73.080.

Chinese classification (CCS): D52.

Estimated length of the Chinese original: 11,144 words.

Standards quoted by this document

GB/T 2007.2; GB/T 6682; GB/T 8170; GB/T 12806; GB/T 12807; GB/T 12808; GB/T 23942