

ICS 13.030.20

CCS H 12

YS

**NONFERROUS METALS INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

YS/T 743-2010

**Determination of gaseous fluoride in purification system for
electrolyzed aluminium by alkali-paper sampling and fluoride
ion-selective electrode analysis**

电解铝净化系统中气氟的测定碱滤纸·氟离子选择性电极法

(English Translation)

Issued on: November 22, 2010

Implemented on: March 1, 2011

Issued by: Ministry of Industry and Information Technology

Table of Contents

What this standard covers	1
Document data	2
Standards quoted by this document	3

What this standard covers

This standard specifies the aluminum purification system gaseous fluoride concentration method. This standard applies to aluminum purification system gaseous fluoride concentration measurement, measurement range: 0. 1 mg/m³ ~ 500 mg/m³. This standard also applies to other sources of gaseous fluoride concentration was determined.

Document data

International classification (ICS): 13.030.20.

Chinese classification (CCS): H12.

Estimated length of the Chinese original: 8,812 words.

Standards quoted by this document

GB/T 16157; GB 25465