

ICS 11.040.30
CCS C 45

YY

**PHARMACEUTICAL INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

YY/T 1493-2016

**Gravity control peritoneal dialysis equipment
(English Translation)**

Issued on: July 29, 2016

Implemented on: June 1, 2018

Issued by: State Food and Drug Administration

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the terms, definitions, classifications, requirements, test methods, signs, instructions for use, packaging, transportation and storage requirements of gravity-controlled peritoneal dialysis equipment. This standard applies to gravity-controlled peritoneal dialysis equipment. This standard does not apply to disposable consumable materials (such as dialysate, dialysate tubing) and pressure-controlled peritoneal dialysis equipment during peritoneal dialysis. This standard does not apply to: disposable consumable materials (such as dialysate, dialysate tubing) during peritoneal dialysis; pressure-controlled peritoneal dialysis equipment.

Document data

International classification (ICS): 11.040.30.

Chinese classification (CCS): C45.

Estimated length of the Chinese original: 10,135 words.