

ICS 11.040.50
CCS C 43

YY

**PHARMACEUTICAL INDUSTRY STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA**

YY/T 1903-2024

**Characteristics and test conditions for cardiac single photon
emission computed tomography device**

心脏单光子发射计算机断层成像装置 性能和试验规则

(English Translation)

Issued on: February 7, 2024

Implemented on: March 1, 2025

Issued by: State Drug Administration

Table of Contents

What this standard covers	1
Document data	2

What this standard covers

This standard specifies the performance of single-photon emission computed tomography (SPECT) systems dedicated to cardiac tomography and describes the corresponding test methods. These systems include SPECT systems dedicated only to cardiac tomography, or general-purpose SPECT systems with a dedicated cardiac tomography subsystem that is outside the scope of the gamma camera standard for planar, whole-body, and SPECT imaging.

Document data

International classification (ICS): 11.040.50.

Chinese classification (CCS): C43.

Estimated length of the Chinese original: 26,215 words.